



**Certificate of Analysis**

**RAPICIDE® PA**

**RTU Part A**

Lot Number: 000318

Expiration Date: 2019/07

Quality Control Report: 180003

Components	Specifications	Analysis
H <sub>2</sub> O <sub>2</sub>	20.0 – 24.0%	<u>21.0 %</u>
PAA	5.0% ± 1.0%	<u>5.6 %</u>
Acetic Acid	8.0 – 10.0%	<u>8.8 %</u>
pH, (as is)	0.5 - 1.1 (as is)	<u>0.9</u>

This lot has been reviewed and meets specified requirements.

\_\_\_\_\_  
Quality Analyst

03 January 2018

\_\_\_\_\_  
Analysis Date

\_\_\_\_\_  
Quality Assurance

\_\_\_\_\_  
Release Date

**Medivators Inc.**  
14605 28th Avenue North  
Minneapolis, MN 55447 USA  
Telephone: (763) 553-3300  
Fax (763) 553-3387  
[www.medivators.com](http://www.medivators.com)