

**Certificate of Analysis**

**RAPICIDE® PA**

**Part A**

Product Code: ML02-0147

Lot Number: 00032218

Expiration Date: 2020/04

Components	Specifications	Analysis
H <sub>2</sub> O <sub>2</sub>	20.0 – 24.0% (w/w)	23.3 %
PAA	4.0 – 6.0 % (w/w)	5.1 %
Acetic Acid	8.0 – 10.0% (w/w)	9.2 %
pH, (as is)	0.5 – 1.1 (as is)	0.8

This lot has been reviewed and meets specified requirements.

  
\_\_\_\_\_  
Quality Analyst

05 Oct 2018  
\_\_\_\_\_  
Analysis Date

  
\_\_\_\_\_  
Quality Assurance

05-Oct-18  
\_\_\_\_\_  
Release Date

**Medivators Inc.**  
14605 28th Avenue North  
Minneapolis, MN 55447 USA  
Telephone: (763) 553-3300  
Fax (763) 553-3387  
[www.medivators.com](http://www.medivators.com)

RAPICIDE® is a registered trademark of Medivators Inc.