

Certificate of Analysis

RAPICIDE[®] PA

Part A

Product Code: **ML02-0147**

Lot Number: **00036019**

Expiration Date: **2021/03**

| Components | Specifications | Analysis |
|-------------------------------|--------------------|---------------|
| H ₂ O ₂ | 20.0 – 24.0% (w/w) | 23.6 % |
| PAA | 4.0 – 6.0 % (w/w) | 5.3 % |
| Acetic Acid | 8.0 – 10.0% (w/w) | 9.0 % |
| pH, (as is) | 0.5 – 1.1 (as is) | 0.9 |

This lot has been reviewed and meets specified requirements.



Quality Analyst



Analysis Date



Quality Assurance



Release Date

Medivators Inc.
14605 28th Avenue North
Minneapolis, MN 55447 USA
Telephone: (763) 553-3300
Fax (763) 553-3387
www.medivators.com

RAPICIDE[®] is a registered trademark of Medivators Inc.