



Certificate of Analysis

INTERCEPT[®] PLUS

Detergent & Cleaner


Part No. ML02-0145


Lot Number: 00039317

Expiration Date: 2019/11

Property	Test Method	Specification	Analysis
Appearance	Visual	Clear Non-viscous Liquid	Pass
pH (release) at 25°C	20002-154	11.5 – 12.5	12.0
Phosphate Level	20602-105	80.83 – 98.79 g/L	85.33 g/L

This lot has been reviewed and meets specified requirements.

 07/12/2017
Quality Analyst Analysis Date

 07-DEC-12
Quality Assurance Release Date

Medivators Inc.
14605 28th Avenue North
Minneapolis, MN 55447 USA
Telephone: (763) 553-3300
Fax (763) 553-3387
www.medivators.com

INTERCEPT[®] is a registered trademark of Medivators Inc.