

**Certificate of Analysis**

**RAPICIDE® PA**

**Part A**

Product Code: **ML02-0143**

Lot Number: 000517

Expiration Date: 2018/07

Components	Specifications	Analysis
H <sub>2</sub> O <sub>2</sub>	20.0 – 24.0% (w/w)	<u>23.5 %</u>
PAA	4.0 – 6.0 % (w/w)	<u>5.2 %</u>
Acetic Acid	8.0 – 10.0% (w/w)	<u>9.0 %</u>
pH, (as is)	0.5 – 1.1 (as is)	<u>0.5</u>

This lot has been reviewed and meets specified requirements.

 13 Jan 2017  
Quality Analyst Analysis Date

 PA 13 Jan 2017  
Quality Assurance Release Date

**Medivators Inc.**  
14605 28th Avenue North  
Minneapolis, MN 55447 USA  
Telephone: (763) 553-3300  
Fax (763) 553-3387  
[www.medivators.com](http://www.medivators.com)

RAPICIDE® is a registered trademark of Medivators Inc.