

Certificate of Analysis

RAPICIDE[®] PA

Part A

Product Code: **ML 02-0115**

Lot Number: 001817

Expiration Date: 2018/07


Components	Specifications	Analysis
H ₂ O ₂	20.0 – 24.0% (w/w)	<u>23.3 %</u>
PAA	4.0 – 6.0 % (w/w)	<u>5.2 %</u>
Acetic Acid	8.0 – 10.0% (w/w)	<u>9.3 %</u>
pH, (as is)	0.5 – 1.1 (as is)	<u>0.6</u>

This lot has been reviewed and meets specified requirements.



Quality Analyst

24 January 2017
Analysis Date



Quality Assurance

24 January 2017
Release Date

Medivators Inc.
14605 28th Avenue North
Minneapolis, MN 55447 USA
Telephone: (763) 553-3300
Fax (763) 553-3387
www.medivators.com

RAPICIDE[®] is a registered trademark of Medivators Inc.