

# Certificate of Analysis

## **RAPICIDE® PA**

### **Part B**


Product Code: **ML02-0115**

Lot Number: **001817**

Expiration Date: **2018/07**

Components	Specification	Analysis
Sodium	743.0 - 909.0 mEq/L	<b>826.3 mEq/L</b>
Benzotriazole	0.9 – 1.1% w/w	<b>1.0 %w/w</b>
pH	11.4 to 12.4 as is	<b>11.6</b>

This lot has been reviewed and meets specified requirements.

  
\_\_\_\_\_  
Quality Analyst

26 January 2017  
\_\_\_\_\_  
Analysis Date

  
\_\_\_\_\_  
Quality Assurance

26 January 2017  
\_\_\_\_\_  
Release Date

**Medivators Inc.**  
14605 28th Avenue North  
Plymouth, Minnesota 55447  
Telephone: (763) 553-3300  
Fax (763) 553-3387

RAPICIDE® is a registered trademark of Medivators Inc.