

Certificate of Analysis

RAPICIDE® PA

Part B

Product Code: **ML02-0115**

Lot Number: **001819**

Expiration Date: **2020/07**

Components	Specification	Analysis
Sodium	743.0 - 909.0 mEq/L	791.0 mEq/L
Benzotriazole	0.9 – 1.1% w/w	1.0 %w/w
pH	11.4 to 12.4 as is	11.9


This lot has been reviewed and meets specified requirements.



Quality Analyst

21-JAN-19

Analysis Date



Quality Assurance

26-JAN-19

Release Date

Medivators Inc.

14605 28th Avenue North
Plymouth, Minnesota 55447
Telephone: (763) 553-3300
Fax (763) 553-3387

RAPICIDE® is a registered trademark of Medivators Inc.