



Certificate of Analysis

RAPICIDE® PA

Part A

Product Code: ML02-0158
Lot Number: 001919
Expiration Date: 2020/07

Components	Specifications	Analysis
H ₂ O ₂	20.0 – 24.0% (w/w)	23.9 %
PAA	4.0 – 6.0 % (w/w)	5.2 %
Acetic Acid	8.0 – 10.0% (w/w)	9.1 %
pH, (as is)	0.5 – 1.1 (as is)	0.6

This lot has been reviewed and meets specified requirements.

PS

Quality Analyst

21/01/19

Analysis Date

Quality Assurance

23-Jan-19

Release Date

Medivators Inc.
14605 28th Avenue North
Minneapolis, MN 55447 USA
Telephone: (763) 553-3300
Fax (763) 553-3387
www.medivators.com

RAPICIDE® is a registered trademark of Medivators Inc.