



Quality Control Report and Certificate of Analysis

Formula: SB-127

Lot #: 001D07

	<u>Specification</u>	<u>Analysis</u>	<u>Deviation</u>
Sodium	<u>4500</u> mEq/L	<u>4490.5</u> mEq/L	<u>-0.21%</u>
Calcium	<u>135</u> mEq/L	<u>132.1</u> mEq/L	<u>-2.11%</u>
Potassium	<u>135</u> mEq/L	<u>133.3</u> mEq/L	<u>-1.22%</u>
Magnesium	<u>33.75</u> mEq/L	<u>32.7</u> mEq/L	<u>-3.26%</u>
Chloride	<u>4803.75</u> mEq/L	<u>4720.9</u> mEq/L	<u>-1.72%</u>
Dextrose	<u>9000</u> mg/dL	<u>9000.0</u> mg/dL	<u>0.00%</u>
Conductivity 25 C	<u>173</u> mS/cm	<u>170.2</u> mS/cm	<u>-1.62%</u>
Pyrogen Test	<u>Non-pyrogenic</u>	<u>Non-pyrogenic</u>	
pH Conc.	<u>1.8 - 3</u> pH Units	<u>2.05</u> pH units	
pH 1:34	<u>3 - 4</u> pH Units	<u>3.49</u> pH units	

Approved for Release:

20600-379/D

Jennifer Ostreich
Quality Control

4/3/07
Date