



Quality Control Report and Certificate of Analysis

Formula: SB-152

Lot #: 001E08

	<u>Specification</u>	<u>Analysis</u>	<u>Deviation</u>
Sodium	<u>4500</u> mEq/L	<u>4499.3</u> mEq/L	<u>-0.02%</u>
Calcium	<u>112.5</u> mEq/L	<u>112.4</u> mEq/L	<u>-0.09%</u>
Potassium	<u>90</u> mEq/L	<u>88.5</u> mEq/L	<u>-1.61%</u>
Magnesium	<u>33.75</u> mEq/L	<u>33.0</u> mEq/L	<u>-2.37%</u>
Chloride	<u>4736.25</u> mEq/L	<u>4720.9</u> mEq/L	<u>-0.32%</u>
Dextrose	<u>9000</u> mg/dL	<u>9000.0</u> mg/dL	<u>0.00%</u>
Conductivity 25 C	<u>164</u> mS/cm	<u>162.9</u> mS/cm	<u>-0.67%</u>
Pyrogen Test	<u>Non-pyrogenic</u>	<u>Non-pyrogenic</u>	
pH Conc.	<u>1.8 - 3</u> pH Units	<u>1.95</u> pH units	
pH 1:34	<u>3 - 4</u> pH Units	<u>3.49</u> pH units	

Approved for Release:

20600-379/D


Quality Control

05-05-08
Date