



Quality Control Report and Certificate of Analysis


Formula: SB-140

Lot #: 002F08

	<u>Specification</u>	<u>Analysis</u>	<u>Deviation</u>
Sodium	<u>4500</u> mEq/L	<u>4487.0</u> mEq/L	<u>-0.29%</u>
Calcium	<u>0</u> mEq/L	<u>0.0</u> mEq/L	<u>0.00%</u>
Potassium	<u>90</u> mEq/L	<u>89.6</u> mEq/L	<u>-0.44%</u>
Magnesium	<u>33.75</u> mEq/L	<u>33.4</u> mEq/L	<u>-1.04%</u>
Chloride	<u>4623.75</u> mEq/L	<u>4578.6</u> mEq/L	<u>-0.98%</u>
Dextrose	<u>9000</u> mg/dL	<u>9000.0</u> mg/dL	<u>0.00%</u>
Conductivity 25 C	<u>174</u> mS/cm	<u>172.3</u> mS/cm	<u>-0.98%</u>
Pyrogen Test	<u>Non-pyrogenic</u>	<u>Non-pyrogenic</u>	
pH Conc.	<u>1.8 - 3</u> pH Units	<u>1.97</u> pH units	
pH 1:34	<u>3 - 4</u> pH Units	<u>3.42</u> pH units	

Approved for Release:

20600-379/D


Quality Control

06-03-08
Date