



Certificate of Conformance

RAPICIDE® OPA-28 Test Strips

Part Number:

ML02-0136

Vial Lot Number:

003093A

Expiration Date:

03/2015

Quality Control Report:

Minntech certifies that the test strips indicate a positive or pass reaction with a solution mixed to a concentration of $0.50\% \pm 0.05\%$ OPA.

Minntech certifies that the test strips indicate a negative or fail reaction with a solution mixed to a concentration of $0.35\% \pm 0.05\%$ OPA.

The statistical analysis and approval test procedure is on file and available for review.

This lot has been reviewed and meets specified requirements.

A handwritten signature in blue ink, appearing to read "Samir Wadhwa".

Quality Assurance

10/28/2013

Release Date

MEDIVATORS Inc.
14605 28th Avenue N
Minneapolis, MN 55447
Tel: 800.444.4729
Fax: 763.553.3387
www.medivators.com