



Quality Control Report and Certificate of Analysis


Formula: SB-111

Lot #: 003E08

	<u>Specification</u>	<u>Analysis</u>	<u>Deviation</u>
Sodium	<u>4500</u> mEq/L	<u>4511.5</u> mEq/L	<u>0.26%</u>
Calcium	<u>135</u> mEq/L	<u>135.8</u> mEq/L	<u>0.56%</u>
Potassium	<u>90</u> mEq/L	<u>89.6</u> mEq/L	<u>-0.44%</u>
Magnesium	<u>33.75</u> mEq/L	<u>33.1</u> mEq/L	<u>-1.93%</u>
Chloride	<u>4758.75</u> mEq/L	<u>4711.6</u> mEq/L	<u>-0.99%</u>
Dextrose	<u>9000</u> mg/dL	<u>9000.0</u> mg/dL	<u>0.00%</u>
Conductivity 25 C	<u>172</u> mS/cm	<u>171.1</u> mS/cm	<u>-0.52%</u>
Pyrogen Test	<u>Non-pyrogenic</u>	<u>Non-pyrogenic</u>	
pH Conc.	<u>1.8 - 3</u> pH Units	<u>1.99</u> pH units	
pH 1:34	<u>3 - 4</u> pH Units	<u>3.51</u> pH units	

Approved for Release:

20600-379/D


Quality Control

05-05-08
Date