



## *Certificate of Conformance*

# RAPICIDE® OPA-28 Test Strips

**Part Number:** **ML02-0136**

**Lot Number:** **004102**

**Expiration Date:** **2014/04**

### **Quality Control Report:**

Minntech certifies that the test strips indicate a positive or pass reaction with a solution mixed to a concentration of 0.50%  $\pm$  0.05% OPA.

Minntech certifies that the test strips indicate a negative or fail reaction with a solution mixed to a concentration of 0.35%  $\pm$  0.05% OPA.

The statistical analysis and approval test procedure is on file and available for review.

This lot has been reviewed and meets specified requirements.

Diane Riskin

Quality Assurance

January 15, 2013

Release Date

**MEDIVATORS Inc.**  
14605 28<sup>th</sup> Avenue N  
Minneapolis, MN 55447  
Tel: 800.444.4729  
Fax: 763.553.3387  
[www.medivators.com](http://www.medivators.com)