

Certificate of Analysis

RAPICIDE[®] PA

Part A

Product Code: **ML02-115**

Lot Number: **004817**

Expiration Date: **2018/08**

Components	Specifications	Analysis
H ₂ O ₂	20.0 – 24.0% (w/w)	23.2 %
PAA	4.0 – 6.0 % (w/w)	5.2 %
Acetic Acid	8.0 – 10.0% (w/w)	9.2 %
pH, (as is)	0.5 – 1.1 (as is)	0.5

This lot has been reviewed and meets specified requirements.

Quality Analyst

28 Febr 2017
Analysis Date

Quality Assurance *PA*

02 March 2017
Release Date

Medivators Inc.
14605 28th Avenue North
Minneapolis, MN 55447 USA
Telephone: (763) 553-3300
Fax (763) 553-3387
www.medivators.com

RAPICIDE[®] is a registered trademark of Medivators Inc.