



Quality Control Report and Certificate of Analysis

Formula: SB-152

Lot #: 004E07

	<u>Specification</u>	<u>Analysis</u>	<u>Deviation</u>
Sodium	<u>4500</u> mEq/L	<u>4485.2</u> mEq/L	<u>-0.33%</u>
Calcium	<u>112.5</u> mEq/L	<u>113.3</u> mEq/L	<u>0.71%</u>
Potassium	<u>90</u> mEq/L	<u>88.9</u> mEq/L	<u>-1.22%</u>
Magnesium	<u>33.75</u> mEq/L	<u>33.5</u> mEq/L	<u>-0.74%</u>
Chloride	<u>4736.25</u> mEq/L	<u>4758.2</u> mEq/L	<u>0.46%</u>
Dextrose	<u>9000</u> mg/dL	<u>9000.0</u> mg/dL	<u>0.00%</u>
Conductivity 25 C	<u>164</u> mS/cm	<u>166.3</u> mS/cm	<u>1.40%</u>
Pyrogen Test	<u>Non-pyrogenic</u>	<u>Non-pyrogenic</u>	
pH Conc.	<u>1.8 - 3</u> pH Units	<u>2.03</u> pH units	
pH 1:34	<u>3 - 4</u> pH Units	<u>3.48</u> pH units	

Approved for Release:

 5/2/07

20600-379/D

Quality Control

Date