



## Certificate of Analysis

# RAPICIDE® PA

### Part A

Product Code:

**ML02-0115**

Lot Number:

005017

Expiration Date:

2018/08

Components	Specifications	Analysis
H <sub>2</sub> O <sub>2</sub>	20.0 – 24.0% (w/w)	23.2 %
PAA	4.0 – 6.0 % (w/w)	5.2 %
Acetic Acid	8.0 – 10.0% (w/w)	9.2 %
pH, (as is)	0.5 – 1.1 (as is)	0.5

This lot has been reviewed and meets specified requirements.

  
Quality Analyst

01 March 2017  
Analysis Date

  
Quality Assurance PA

03 March 2017  
Release Date

Medivators Inc.  
14605 28th Avenue North  
Minneapolis, MN 55447 USA  
Telephone: (763) 553-3300  
Fax (763) 553-3387  
[www.medivators.com](http://www.medivators.com)

RAPICIDE® is a registered trademark of Medivators Inc.