



Quality Control Report and Certificate of Analysis

Formula: MB-330-L

Lot #: 005C07

	<u>Specification</u>	<u>Analysis</u>	<u>Deviation</u>
Sodium	<u>967.2</u> mEq/L	<u>953.5</u> mEq/L	<u>-1.42%</u>
Total Bicarbonate and Carbonate	<u>967.2</u> mEq/L	<u>980.0</u> mEq/L	<u>1.32%</u>
Hydrogen peroxide	<u>0</u> ppm	<u>0</u> ppm	
Conductivity 25 C	<u>45</u> mS/cm	<u>45</u> mS/cm	<u>0</u> mS/cm
pH Conc.	<u>7 - 8.3</u> pH Units	<u>8.03</u> pH units	
Pyrogen Test	<u>Non-pyrogenic</u>	<u>Non-pyrogenic</u>	
Bacteriological Tests:			
Total Aerobic Plate Count	<u>< 1.0 CFU/200mL filtered</u>	<u>NG</u> CFU/200 mL*	
Yeast and Mold Tests	<u>< 1.0 CFU/200mL filtered</u>	<u>NG</u> CFU/200 mL*	

*NG = No Growth Observed During Testing

Expiration Date: 18 months from production date.

Release Date 9-12 days from production date if specifications are met.

Note: LAL Pyrogen Sensitivity: 0.06 EU/mL; 1 EU = 1/10 ng; 1 ng = 10 EU; Sensitivity = 0.006ng/mL.

Approved for Release:

20600-379/D
 3/20/07
Quality Control Date