



Quality Control Report and Certificate of Analysis


Formula: SB-1003

Lot #: 005E08

	<u>Specification</u>	<u>Analysis</u>	<u>Deviation</u>
Sodium	<u>2946.4</u> mEq/L	<u>2936.5</u> mEq/L	<u>-0.34%</u>
Calcium	<u>128.9</u> mEq/L	<u>126.2</u> mEq/L	<u>-2.09%</u>
Potassium	<u>73.7</u> mEq/L	<u>73.1</u> mEq/L	<u>-0.75%</u>
Magnesium	<u>36.8</u> mEq/L	<u>36.3</u> mEq/L	<u>-1.49%</u>
Chloride	<u>3185.8</u> mEq/L	<u>3165.8</u> mEq/L	<u>-0.63%</u>
Dextrose	<u>7366</u> mg/dL	<u>7400.0</u> mg/dL	<u>0.46%</u>
Pyrogen Test	<u>Non-pyrogenic</u>	<u>Non-pyrogenic</u>	
pH Conc.	<u>2 - 3</u> pH Units	<u>2.04</u> pH units	
pH 1:34	<u>3 - 4</u> pH Units	<u>3.58</u> pH units	

Approved for Release:

20600-379/D


Quality Control

05-05-08
Date