



Quality Control Report and Certificate of Analysis

Formula: SB-1057

Lot #: 005G07

	<u>Specification</u>	<u>Analysis</u>	<u>Deviation</u>
Sodium	<u>2946.4</u> mEq/L	<u>2927.7</u> mEq/L	<u>-0.63%</u>
Calcium	<u>92.1</u> mEq/L	<u>89.1</u> mEq/L	<u>-3.31%</u>
Potassium	<u>110.5</u> mEq/L	<u>109.9</u> mEq/L	<u>-0.54%</u>
Magnesium	<u>27.6</u> mEq/L	<u>27.5</u> mEq/L	<u>-0.36%</u>
Chloride	<u>3176.6</u> mEq/L	<u>3200.8</u> mEq/L	<u>0.76%</u>
Dextrose	<u>7366</u> mg/dL	<u>7100.0</u> mg/dL	<u>-3.61%</u>
Conductivity 25 C	<u>150.8</u> mS/cm	<u>150.1</u> mS/cm	<u>-0.46%</u>
Pyrogen Test	<u>Non-pyrogenic</u>	<u>Non-pyrogenic</u>	
pH Conc.	<u>2 - 3</u> pH Units	<u>2.26</u> pH units	
pH 1:34	<u>3 - 4</u> pH Units	<u>3.53</u> pH units	

Approved for Release:

20600-379/D

John Doe
Quality Control

07-03-09
Date