



## Quality Control Report and Certificate of Analysis

Formula: MB-330-L

Lot #: 005K07

	<u>Specification</u>	<u>Analysis</u>	<u>Deviation</u>
Sodium	<u>967.2 mEq/L</u>	<u>951.5 mEq/L</u>	<u>-1.62%</u>
Total Bicarbonate and Carbonate	<u>967.2 mEq/L</u>	<u>960.0 mEq/L</u>	<u>-0.74%</u>
Hydrogen peroxide	<u>0 ppm</u>	<u>0 ppm</u>	
Conductivity 25 C	<u>45 mS/cm</u>	<u>46 mS/cm</u>	<u>1 mS/cm</u>
pH Conc.	<u>7 - 8.3 pH Units</u>	<u>8.11 pH units</u>	
Pyrogen Test	<u>Non-pyrogenic</u>	<u>Non-pyrogenic</u>	
Bacteriological Tests:			
Total Aerobic Plate Count	<u>&lt; 1.0 CFU/200mL filtered</u>	<u>NG CFU/200 mL*</u>	
Yeast and Mold Tests	<u>&lt; 1.0 CFU/200mL filtered</u>	<u>NG CFU/200 mL*</u>	

\*NG = No Growth Observed During Testing

Expiration Date: 18 months from production date.

Release Date 9-12 days from production date if specifications are met.

Note: LAL Pyrogen Sensitivity: 0.06 EU/mL; 1 EU = 1/10 ng; 1 ng = 10 EU; Sensitivity = 0.006ng/mL.

Approved for Release:

Elsie Troy  
Quality Control

10/17/07

Date