

Certificate of Conformance

MaxiCide® OPA 28 Test Strips

Part Number:

112-6977

Kit Lot Number:

006085

Vial Lot Number:

026085

Expiration Date:

04/17

Quality Control Report:

MEDIVATORS certifies that the test strips indicate a positive or pass reaction with a solution mixed to a concentration of 0.50% \pm 0.05% OPA.

MEDIVATORS certifies that the test strips indicate a negative or fail reaction with a solution mixed to a concentration of 0.35% \pm 0.05% OPA.

The statistical analysis and approval test procedure is on file and available for review.

This lot has been reviewed and meets specified requirements.



Linda Heidtken

Quality Assurance

11/23/2015

Release Date

Medivators Inc.

14605 28th Avenue N

Minneapolis, MN 55447

Tel: 800.444.4729

Fax: 763.553.3387

www.medivators.com