



Certificate of Analysis

RAPICIDE® PA


Part B

Product Code: ML02-0115
Lot Number: 007219
Expiration Date: 2020/08


Components	Specification	Analysis
Sodium	743.0 - 909.0 mEq/L	780.0 mEq/L
Benzotriazole	0.9 – 1.1% w/w	1.1 %w/w
pH	11.4 to 12.4 as is	11.5

This lot has been reviewed and meets specified requirements.


Quality Analyst


25/02/13
Analysis Date

Quality Assurance


27 FEB-13
Release Date
Medivators Inc.
14605 28th Avenue North
Plymouth, Minnesota 55447
Telephone: (763) 553-3300
Fax (763) 553-3387

RAPICIDE® is a registered trademark of Medivators Inc.