



Certificate of Analysis

RAPICIDE™ PA

RTU Part A

Lot Number: 007519
Expiration Date: 2020/08

Components	Specifications	Analysis
H ₂ O ₂	20.0 – 24.0%	23.3%
PAA	5.0% \pm 1.0%	5.5%
Acetic Acid	8.0 – 10.0%	8.9%
pH, (as is)	0.5 - 1.1 (as is)	0.5

This lot has been reviewed and meets specified requirements.

FDC

Quality Analyst

Quality Assurance

27-FEB-19

Analysis Date

01-MAR-19

Release Date

Medivators Inc.

14605 28th Avenue North
Minneapolis, MN 55447 USA
Telephone: (763) 553-3300
Fax (763) 553-3387
www.medivators.com

RAPICIDE™ is a trademark of Medivators Inc.