



## Certificate of Analysis

# RAPICIDE™ PA

## RTU Part B

Lot Number: 007619  
Expiration Date: 2020/08

Components	Specifications	Analysis
Na+	35.5 – 43.4 mEq/L	<u>36.0mEq/L</u>
Benzotriazole	0.05% $\pm$ 0.005%	<u>0.046%</u>
pH, (as is)	10.6 – 11.6 (as is)	<u>11.6</u>

This lot has been reviewed and meets specified requirements.

RE  
Quality Analyst

27-FEB-13  
Analysis Date

CC  
Quality Assurance

01-MAR-13  
Release Date

**Medivators Inc.**  
14605 28th Avenue North  
Minneapolis, MN 55447 USA  
Telephone: (763) 553-3300  
Fax (763) 553-3387  
[www.medivators.com](http://www.medivators.com)

RAPICIDE™ is a trademark of Medivators Inc.