



Certificate of Analysis

RAPICIDE® PA

Part B

Product Code:

ML02-0115

Lot Number:

008316

Expiration Date:

10/2017

Components	Specification	Analysis
Sodium	743.0 - 909.0 mEq/L	805.7 mEq/L
Benzotriazole	0.9 – 1.1%	1.0%
pH	11.4 to 12.4 as is	11.7

This lot has been reviewed and meets specified requirements.

Quality Analyst

02MAY2016

Analysis Date

Quality Assurance

02MAY2016

Release Date

Medivators Inc.

14605 28th Avenue North
Plymouth, Minnesota 55447
Telephone: (763) 553-3300
Fax (763) 553-3387

RAPICIDE® is a registered trademark of Medivators Inc.