



## Certificate of Analysis

# RAPICIDE® PA


### Part A

Product Code: **ML02-0115**  
Lot Number: 008416  
Expiration Date: 2017/10

Components	Specifications	Analysis
H <sub>2</sub> O <sub>2</sub>	20.0 – 24.0% (w/w)	23.1 %
PAA	4.0 – 6.0 % (w/w)	5.7 %
Acetic Acid	8.0 – 10.0% (w/w)	8.9 %
pH, (as is)	0.5 – 1.1 (as is)	0.8

This lot has been reviewed and meets specified requirements.

  
\_\_\_\_\_  
Quality Analyst 14April2016  
Analysis Date

  
\_\_\_\_\_  
Quality Assurance 14April2016  
Release Date

**Medivators Inc.**  
14605 28th Avenue North  
Minneapolis, MN 55447 USA  
Telephone: (763) 553-3300  
Fax (763) 553-3387  
[www.medivators.com](http://www.medivators.com)

RAPICIDE® is a registered trademark of Medivators Inc.