



## Certificate of Analysis

# RAPICIDE™ PA

## RTU Part B

Lot Number: 008619  
Expiration Date: 2020/08

Components	Specifications	Analysis
Na+	35.5 – 43.4 mEq/L	36.2mEq/L
Benzotriazole	0.05% $\pm$ 0.005%	0.047%
pH, (as is)	10.6 – 11.6 (as is)	11.6

This lot has been reviewed and meets specified requirements.

RE  
Quality Analyst

06-MAR-19  
Analysis Date

QC  
Quality Assurance

06-MAR-19  
Release Date

Medivators Inc.  
14605 28th Avenue North  
Minneapolis, MN 55447 USA  
Telephone: (763) 553-3300  
Fax (763) 553-3387  
[www.medivators.com](http://www.medivators.com)

RAPICIDE™ is a trademark of Medivators Inc.