



Certificate of Analysis

RAPICIDE™ PA

RTU Part B

Lot Number: 008818
Expiration Date: 2019/08

Components	Specifications	Analysis
Na+	35.5 – 43.4 mEq/L	37.1mEq/L
Benzotriazole	0.05% \pm 0.005%	0.053%
pH, (as is)	10.6 – 11.6 (as is)	11.6

This lot has been reviewed and meets specified requirements.

PDC

Quality Analyst

28 Feb' 18

Analysis Date

pp Jank G

Quality Assurance

28 March '18

Release Date

Medivators Inc.
14605 28th Avenue North
Minneapolis, MN 55447 USA
Telephone: (763) 553-3300
Fax (763) 553-3387
www.medivators.com

RAPICIDE™ is a trademark of Medivators Inc.