



Quality Control Report and Certificate of Analysis

Formula: SB-1052

Lot #: 008B08

	<u>Specification</u>	<u>Analysis</u>	<u>Deviation</u>
Sodium	<u>3093.7</u> mEq/L	<u>3039.8</u> mEq/L	<u>-1.74%</u>
Calcium	<u>119.7</u> mEq/L	<u>118.7</u> mEq/L	<u>-0.88%</u>
Potassium	<u>73.7</u> mEq/L	<u>73.7</u> mEq/L	<u>-0.03%</u>
Magnesium	<u>36.8</u> mEq/L	<u>36.9</u> mEq/L	<u>0.27%</u>
Chloride	<u>3323.9</u> mEq/L	<u>3306.9</u> mEq/L	<u>-0.51%</u>
Dextrose	<u>7366</u> mg/dL	<u>7400.0</u> mg/dL	<u>0.46%</u>
Conductivity 25 C	<u>154</u> mS/cm	<u>155.2</u> mS/cm	<u>0.78%</u>
Pyrogen Test	<u>Non-pyrogenic</u>	<u>Non-pyrogenic</u>	
pH Conc.	<u>2 - 3</u> pH Units	<u>2.08</u> pH units	
pH 1:34	<u>3 - 4</u> pH Units	<u>3.59</u> pH units	

Approved for Release:

Quality Control

02-01-08
Date