




Quality Control Report and Certificate of Analysis

Formula: SB-1020

Lot #: 008E07

	<u>Specification</u>	<u>Analysis</u>	<u>Deviation</u>
Sodium	<u>2946.4</u> mEq/L	<u>2896.3</u> mEq/L	<u>-1.70%</u>
Calcium	<u>92.1</u> mEq/L	<u>93.6</u> mEq/L	<u>1.63%</u>
Potassium	<u>73.7</u> mEq/L	<u>73.1</u> mEq/L	<u>-0.75%</u>
Magnesium	<u>36.8</u> mEq/L	<u>38.0</u> mEq/L	<u>3.13%</u>
Chloride	<u>3149</u> mEq/L	<u>3155.3</u> mEq/L	<u>0.20%</u>
Dextrose	<u>7366</u> mg/dL	<u>7300.0</u> mg/dL	<u>-0.90%</u>
Conductivity 25 C	<u>148</u> mS/cm	<u>149.3</u> mS/cm	<u>0.88%</u>
Pyrogen Test	<u>Non-pyrogenic</u>	<u>Non-pyrogenic</u>	
pH Conc.	<u>2 - 3</u> pH Units	<u>2.26</u> pH units	
pH 1:34	<u>3 - 4</u> pH Units	<u>3.42</u> pH units	

Approved for Release:



Quality Control

5/2/02
Date