



Quality Control Report and Certificate of Analysis


Formula: SB-142

Lot #: 008E08

	<u>Specification</u>	<u>Analysis</u>	<u>Deviation</u>
Sodium	<u>4500</u> mEq/L	<u>4499.3</u> mEq/L	<u>-0.02%</u>
Calcium	<u>101.25</u> mEq/L	<u>100.9</u> mEq/L	<u>-0.35%</u>
Potassium	<u>45</u> mEq/L	<u>45.3</u> mEq/L	<u>0.72%</u>
Magnesium	<u>33.75</u> mEq/L	<u>33.3</u> mEq/L	<u>-1.33%</u>
Chloride	<u>4680</u> mEq/L	<u>4681.8</u> mEq/L	<u>0.04%</u>
Dextrose	<u>9000</u> mg/dL	<u>9000.0</u> mg/dL	<u>0.00%</u>
Conductivity 25 C	<u>170</u> mS/cm	<u>168.8</u> mS/cm	<u>-0.71%</u>
Pyrogen Test	<u>Non-pyrogenic</u>	<u>Non-pyrogenic</u>	
pH Conc.	<u>1.8 - 3</u> pH Units	<u>1.99</u> pH units	
pH 1:34	<u>3 - 4</u> pH Units	<u>3.52</u> pH units	

Approved for Release:

20600-379/D


Quality Control

05-01-08
Date