



Quality Control Report and Certificate of Analysis


Formula: SB-1078

Lot #: 008G08

	<u>Specification</u>	<u>Analysis</u>	<u>Deviation</u>
Sodium	<u>2983.2</u> mEq/L	<u>2962.8</u> mEq/L	<u>-0.69%</u>
Calcium	<u>92.1</u> mEq/L	<u>90.3</u> mEq/L	<u>-2.01%</u>
Potassium	<u>36.8</u> mEq/L	<u>37.5</u> mEq/L	<u>1.77%</u>
Magnesium	<u>36.8</u> mEq/L	<u>36.3</u> mEq/L	<u>-1.49%</u>
Chloride	<u>3149</u> mEq/L	<u>3120.3</u> mEq/L	<u>-0.91%</u>
Dextrose	<u>7366</u> mg/dL	<u>7400.0</u> mg/dL	<u>0.46%</u>
Conductivity 25 C	<u>151</u> mS/cm	<u>150.5</u> mS/cm	<u>-0.33%</u>
Pyrogen Test	<u>Non-pyrogenic</u>	<u>Non-pyrogenic</u>	
pH Conc.	<u>2 - 3</u> pH Units	<u>2.1</u> pH units	
pH 1:34	<u>3 - 4</u> pH Units	<u>3.62</u> pH units	

Approved for Release:

20600-379/D


Quality Control

07-02-08
Date