

Certificate of Conformance

MaxiCide® OPA 28 Test Strips

Part Number:	112-6977
Vial Lot Number:	009044
Expiration Date:	10/15

Quality Control Report:

MEDIVATORS certifies that the test strips indicate a positive or pass reaction with a solution mixed to a concentration of 0.50% \pm 0.05% OPA.

MEDIVATORS certifies that the test strips indicate a negative or fail reaction with a solution mixed to a concentration of 0.35% \pm 0.05% OPA.

The statistical analysis and approval test procedure is on file and available for review.

This lot has been reviewed and meets specified requirements.



T. B. Jones
Quality Assurance

6/9/2014
Release Date

Medivators Inc.
14605 28th Avenue N
Minneapolis, MN 55447
Tel: 800.444.4729
Fax: 763.553.3387
www.medivators.com