



Certificate of Analysis

RAPICIDE™ PA

RTU Part B

Lot Number: 009218
Expiration Date: 2019/08

Components	Specifications	Analysis
Na+	35.5 – 43.4 mEq/L	36.6mEq/L
Benzotriazole	0.05% \pm 0.005%	0.055%
pH, (as is)	10.6 – 11.6 (as is)	11.6

This lot has been reviewed and meets specified requirements.

EDC

Quality Analyst 28 Feb' 18
Analysis Date

pp frank Jr

Quality Assurance 28 Feb' 18
Release Date

Medivators Inc.
14605 28th Avenue North
Minneapolis, MN 55447 USA
Telephone: (763) 553-3300
Fax (763) 553-3387
www.medivators.com

RAPICIDE™ is a trademark of Medivators Inc.