



## Certificate of Analysis

# RAPICIDE® PA

### Part A

Product Code:

**ML02-115**

Lot Number:

009818

Expiration Date:

2019/09

Components	Specifications	Analysis
H <sub>2</sub> O <sub>2</sub>	20.0 – 24.0% (w/w)	22.0 %
PAA	4.0 – 6.0 % (w/w)	5.1 %
Acetic Acid	8.0 – 10.0% (w/w)	9.2 %
pH, (as is)	0.5 – 1.1 (as is)	0.7

This lot has been reviewed and meets specified requirements.

Quality Analyst

07 March 2018

Analysis Date

07 March '18

Release Date

Medivators Inc.  
14605 28th Avenue North  
Minneapolis, MN 55447 USA  
Telephone: (763) 553-3300  
Fax (763) 553-3387  
www.medivators.com

RAPICIDE® is a registered trademark of Medivators Inc.