


Quality Control Report and Certificate of Analysis

Formula: SB-111Lot #: 009C07

	<u>Specification</u>	<u>Analysis</u>	<u>Deviation</u>
Sodium	<u>4500</u> mEq/L	<u>4532.5</u> mEq/L	<u>0.72%</u>
Calcium	<u>135</u> mEq/L	<u>131.5</u> mEq/L	<u>-2.59%</u>
Potassium	<u>90</u> mEq/L	<u>88.4</u> mEq/L	<u>-1.81%</u>
Magnesium	<u>33.75</u> mEq/L	<u>33.4</u> mEq/L	<u>-1.04%</u>
Chloride	<u>4758.75</u> mEq/L	<u>4795.6</u> mEq/L	<u>0.77%</u>
Dextrose	<u>9000</u> mg/dL	<u>9100.0</u> mg/dL	<u>1.11%</u>
Conductivity 25 C	<u>172</u> mS/cm	<u>171.0</u> mS/cm	<u>-0.58%</u>
Pyrogen Test	<u>Non-pyrogenic</u>	<u>Non-pyrogenic</u>	
pH Conc.	<u>1.8 - 3</u> pH Units	<u>2.13</u> pH units	
pH 1:34	<u>3 - 4</u> pH Units	<u>3.5</u> pH units	

Approved for Release:

20600-379/D



Quality Control3-2-07

Date