



Quality Control Report and Certificate of Analysis

Formula: SB-151

Lot #: 009D07

	<u>Specification</u>	<u>Analysis</u>	<u>Deviation</u>
Sodium	<u>4500</u> mEq/L	<u>4536.0</u> mEq/L	<u>0.80%</u>
Calcium	<u>90</u> mEq/L	<u>91.2</u> mEq/L	<u>1.33%</u>
Potassium	<u>90</u> mEq/L	<u>90.1</u> mEq/L	<u>0.14%</u>
Magnesium	<u>33.75</u> mEq/L	<u>33.1</u> mEq/L	<u>-1.93%</u>
Chloride	<u>4713.75</u> mEq/L	<u>4742.5</u> mEq/L	<u>0.61%</u>
Dextrose	<u>9000</u> mg/dL	<u>9000.0</u> mg/dL	<u>0.00%</u>
Conductivity 25 C	<u>168</u> mS/cm	<u>169.3</u> mS/cm	<u>0.77%</u>
Pyrogen Test	<u>Non-pyrogenic</u>	<u>Non-pyrogenic</u>	
pH Conc.	<u>1.8 - 3</u> pH Units	<u>1.98</u> pH units	
pH 1:34	<u>3 - 4</u> pH Units	<u>3.4</u> pH units	

Approved for Release:

Jennifer Destrich 4/3/07
Quality Control Date