



Quality Control Report and Certificate of Analysis

Formula: SB-152

Lot #: 009E07

	<u>Specification</u>	<u>Analysis</u>	<u>Deviation</u>
Sodium	<u>4500</u> mEq/L	<u>4480.0</u> mEq/L	<u>-0.44%</u>
Calcium	<u>112.5</u> mEq/L	<u>113.9</u> mEq/L	<u>1.24%</u>
Potassium	<u>90</u> mEq/L	<u>90.1</u> mEq/L	<u>0.14%</u>
Magnesium	<u>33.75</u> mEq/L	<u>34.5</u> mEq/L	<u>2.22%</u>
Chloride	<u>4736.25</u> mEq/L	<u>4725.0</u> mEq/L	<u>-0.24%</u>
Dextrose	<u>9000</u> mg/dL	<u>9000.0</u> mg/dL	<u>0.00%</u>
Conductivity 25 C	<u>164</u> mS/cm	<u>169.4</u> mS/cm	<u>3.29%</u>
Pyrogen Test	<u>Non-pyrogenic</u>	<u>Non-pyrogenic</u>	
pH Conc.	<u>1.8 - 3</u> pH Units	<u>2.01</u> pH units	
pH 1:34	<u>3 - 4</u> pH Units	<u>3.31</u> pH units	

Approved for Release:


Quality Control

5/2/07
Date