



## Quality Control Report and Certificate of Analysis


Formula: SB-1020

Lot #: 009E08

|                   | <u>Specification</u>  | <u>Analysis</u>      | <u>Deviation</u> |
|-------------------|-----------------------|----------------------|------------------|
| Sodium            | <u>2946.4</u> mEq/L   | <u>2905.0</u> mEq/L  | <u>-1.41%</u>    |
| Calcium           | <u>92.1</u> mEq/L     | <u>92.6</u> mEq/L    | <u>0.49%</u>     |
| Potassium         | <u>73.7</u> mEq/L     | <u>73.5</u> mEq/L    | <u>-0.27%</u>    |
| Magnesium         | <u>36.8</u> mEq/L     | <u>35.9</u> mEq/L    | <u>-2.45%</u>    |
| Chloride          | <u>3149</u> mEq/L     | <u>3137.2</u> mEq/L  | <u>-0.38%</u>    |
| Dextrose          | <u>7366</u> mg/dL     | <u>7400.0</u> mg/dL  | <u>0.46%</u>     |
| Conductivity 25 C | <u>148</u> mS/cm      | <u>147.3</u> mS/cm   | <u>-0.47%</u>    |
| Pyrogen Test      | <u>Non-pyrogenic</u>  | <u>Non-pyrogenic</u> |                  |
| pH Conc.          | <u>2 - 3</u> pH Units | <u>2.08</u> pH units |                  |
| pH 1:34           | <u>3 - 4</u> pH Units | <u>3.61</u> pH units |                  |

Approved for Release:

20600-379/D

  
Quality Control

05-05-08  
Date