



Quality Control Report and Certificate of Analysis

Formula: SB-144

Lot #: 009F08

	<u>Specification</u>	<u>Analysis</u>	<u>Deviation</u>
Sodium	<u>4500</u> mEq/L	<u>4529.0</u> mEq/L	<u>0.64%</u>
Calcium	<u>101.25</u> mEq/L	<u>100.4</u> mEq/L	<u>-0.84%</u>
Potassium	<u>135</u> mEq/L	<u>135.4</u> mEq/L	<u>0.33%</u>
Magnesium	<u>33.75</u> mEq/L	<u>33.6</u> mEq/L	<u>-0.44%</u>
Chloride	<u>4770</u> mEq/L	<u>4746.0</u> mEq/L	<u>-0.50%</u>
Dextrose	<u>9000</u> mg/dL	<u>9000.0</u> mg/dL	<u>0.00%</u>
Conductivity 25 C	<u>172.5</u> mS/cm	<u>171.1</u> mS/cm	<u>-0.81%</u>
Pyrogen Test	<u>Non-pyrogenic</u>	<u>Non-pyrogenic</u>	
pH Conc.	<u>1.8 - 3</u> pH Units	<u>1.96</u> pH units	
pH 1:34	<u>3 - 4</u> pH Units	<u>3.46</u> pH units	

Approved for Release:

20600-379/D


Quality Control

06-03-08
Date