



Quality Control Report and Certificate of Analysis

Formula: SB-144

Lot #: 009G07

	<u>Specification</u>	<u>Analysis</u>	<u>Deviation</u>
Sodium	<u>4500</u> mEq/L	<u>4525.5</u> mEq/L	<u>0.57%</u>
Calcium	<u>101.25</u> mEq/L	<u>103.8</u> mEq/L	<u>2.52%</u>
Potassium	<u>135</u> mEq/L	<u>136.1</u> mEq/L	<u>0.85%</u>
Magnesium	<u>33.75</u> mEq/L	<u>33.5</u> mEq/L	<u>-0.59%</u>
Chloride	<u>4770</u> mEq/L	<u>4863.8</u> mEq/L	<u>1.97%</u>
Dextrose	<u>9000</u> mg/dL	<u>9000.0</u> mg/dL	<u>0.00%</u>
Conductivity 25 C	<u>172.5</u> mS/cm	<u>170.3</u> mS/cm	<u>-1.28%</u>
Pyrogen Test	<u>Non-pyrogenic</u>	<u>Non-pyrogenic</u>	
pH Conc.	<u>1.8 - 3</u> pH Units	<u>2.16</u> pH units	
pH 1:34	<u>3 - 4</u> pH Units	<u>3.48</u> pH units	

Approved for Release:

20600-379/D

Zachary D. V. C.
Quality Control

07-03-07
Date