



Certificate of Analysis

RAPICIDE® PA

Part A

Product Code: ML 02-0115
Lot Number: 010517
Expiration Date: 2018/10

Components	Specifications	Analysis
H ₂ O ₂	20.0 – 24.0% (w/w)	23.3 %
PAA	4.0 – 6.0 % (w/w)	5.2 %
Acetic Acid	8.0 – 10.0% (w/w)	9.2 %
pH, (as is)	0.5 – 1.1 (as is)	0.5

This lot has been reviewed and meets specified requirements.

Quality Analyst

26 April 2017

Analysis Date

Quality Assurance

02 May 2017

Release Date

Medivators Inc.

14605 28th Avenue North

Minneapolis, MN 55447 USA

Telephone: (763) 553-3300

Fax (763) 553-3387

www.medivators.com

RAPICIDE® is a registered trademark of Medivators Inc.