



Certificate of Analysis

RAPICIDE® PA

Part A

Product Code: ML02-0115
Lot Number: 010916
Expiration Date: 2017/11


Components	Specifications	Analysis
H ₂ O ₂	20.0 – 24.0% (w/w)	23.4 %
PAA	4.0 – 6.0 % (w/w)	5.6 %
Acetic Acid	8.0 – 10.0% (w/w)	8.1 %
pH, (as is)	0.5 – 1.1 (as is)	0.8

This lot has been reviewed and meets specified requirements.



Quality Analyst

12May2016
Analysis Date



Quality Assurance
Q. P.

12May2016
Release Date

Medivators Inc.
14605 28th Avenue North
Minneapolis, MN 55447 USA
Telephone: (763) 553-3300
Fax (763) 553-3387
www.medivators.com

RAPICIDE® is a registered trademark of Medivators Inc.