

Certificate of Analysis


RAPICIDE® PA

Part B

Product Code: **ML02-0115**
Lot Number: **011116**
Expiration Date: **2017/11**

Components	Test Method	Specification	Analysis
Sodium	20100-070	743.0 - 909.0 mEq/L	808.5 mEq/L
Benzotriazole	20601-489	0.9 – 1.1% w/w	1.0 %w/w
pH	20002-154	11.4 to 12.4 as is	11.7

This lot has been reviewed and meets specified requirements.



Quality Analyst

20May2016
Analysis Date



Quality Assurance

20May2016
Release Date

Medivators Inc.
14605 28th Avenue North
Plymouth, Minnesota 55447
Telephone: (763) 553-3300
Fax (763) 553-3387

RAPICIDE® is a registered trademark of Medivators Inc.