



## *Certificate of Conformance*

# **RAPICIDE® OPA/28 Test Strips**

<b>Part Number:</b>	<b>73-OPA28B</b>
<b>Lot Number:</b>	011124
<b>Vial Lot Number:</b>	017124
<b>Expiration Date:</b>	06/16

### **Quality Control Report:**

MEDIVATORS certifies that the test strips indicate a positive or pass reaction with a solution mixed to a concentration of  $0.50\% \pm 0.05\%$  OPA.

MEDIVATORS certifies that the test strips indicate a negative or fail reaction with a solution mixed to a concentration of  $0.35\% \pm 0.05\%$  OPA.

The statistical analysis and approval test procedure is on file and available for review.

This lot has been reviewed and meets specified requirements.

  
Quality Assurance

1/5/2015  
Release Date

MEDIVATORS Inc.  
14605 28<sup>th</sup> Avenue N  
Minneapolis, MN 55447  
Tel: 800.444.4729  
Fax: 763.553.3387  
[www.medivators.com](http://www.medivators.com)