



## Certificate of Analysis

# RAPICIDE® PA

### Part A

Product Code:

**ML02-0143**

Lot Number:

011217

Expiration Date:

2018/10

Components	Specifications	Analysis
H <sub>2</sub> O <sub>2</sub>	20.0 – 24.0% (w/w)	23.3 %
PAA	4.0 – 6.0 % (w/w)	5.2 %
Acetic Acid	8.0 – 10.0% (w/w)	9.2 %
pH, (as is)	0.5 – 1.1 (as is)	0.5

This lot has been reviewed and meets specified requirements.

Quality Analyst

12 April 2017

Analysis Date

Quality Assurance

12 April 2017

Release Date

Medivators Inc.  
14605 28th Avenue North  
Minneapolis, MN 55447 USA  
Telephone: (763) 553-3300  
Fax (763) 553-3387  
[www.medivators.com](http://www.medivators.com)

RAPICIDE® is a registered trademark of Medivators Inc.