



Quality Control Report and Certificate of Analysis

Formula: BC-1-L

Lot #: 011F08

	<u>Specification</u>	<u>Analysis</u>	<u>Deviation</u>
Sodium	<u>1187.7</u> mEq/L	<u>1196.5</u> mEq/L	<u>0.74%</u>
Chloride	<u>402.6</u> mEq/L	<u>400.8</u> mEq/L	<u>-0.44%</u>
Total Bicarbonate and Carbonate	<u>785.1</u> mEq/L	<u>780.0</u> mEq/L	<u>-0.65%</u>
Hydrogen peroxide	<u>0</u> ppm	<u>0.0</u> ppm	
Conductivity 25 C	<u>67</u> mS/cm	<u>66</u> mS/cm	<u>-1 mS/cm</u>
pH Conc.	<u>7 - 8.3</u> pH Units	<u>8.12</u> pH units	
Pyrogen Test	<u>Non-pyrogenic</u>	<u>Non-pyrogenic</u>	
Bacteriological Tests:			
Total Aerobic Plate Count	<u>< 1.0 CFU/200mL filtered</u>	<u>NG</u> CFU/200 mL*	
Yeast and Mold Tests	<u>< 1.0 CFU/200mL filtered</u>	<u>NG</u> CFU/200 mL*	

*NG = No Growth Observed During Testing.

Expiration Date: 18 months from production date.

Release Date 9-12 days from production date if specifications are met.

Note: LAL Pyrogen Sensitivity: 0.06 EU/mL; 1 EU = 1/10 ng; 1 ng = 10 EU; Sensitivity = 0.006ng/mL.

Approved for Release:

Elaine Troje 06/19/08
Quality Control Date