

# *Certificate of Conformance*

## **McKESSON® OPA/28 Test Strips**

**Part Number:**

**73-OPA28B**

**Vial Lot Number:**

012084

**Expiration Date:**

02/16

### **Quality Control Report:**

MEDIVATORS certifies that the test strips indicate a positive or pass reaction with a solution mixed to a concentration of  $0.50\% \pm 0.05\%$  OPA.

MEDIVATORS certifies that the test strips indicate a negative or fail reaction with a solution mixed to a concentration of  $0.35\% \pm 0.05\%$  OPA.

The statistical analysis and approval test procedure is on file and available for review.

This lot has been reviewed and meets specified requirements.



*Quality Assurance*

9/24/2014

*Release Date*

### **MEDIVATORS Inc.**

14605 28<sup>th</sup> Avenue N

Minneapolis, MN 55447

Tel: 800.444.4729

Fax: 763.553.3387

[www.medivators.com](http://www.medivators.com)